



ILLINOIS ENVIRONMENTAL PROTECTION AGENCY

1021 NORTH GRAND AVENUE EAST, P.O. BOX 19276, SPRINGFIELD, ILLINOIS 62794-9276 • (217) 782-3397

JB PRITZKER, GOVERNOR

JOHN J. KIM, DIRECTOR

Inspection Report

GENERAL INFORMATION			
Report Date:	July 3, 2019	Inspection Date:	June 27, 2019
Inspector:	Daniel Delgado	Last Inspection Date:	June 20, 2019

SOURCE INFORMATION	
LOCATION	
Facility ID #:	043110AAC
Company Name:	Sterigenics
Street Address:	7775 S. Quincy Street
City, County:	Willowbrook, Du Page
State, Zip Code:	Illinois, 60527
Contact/Title:	Paul Krett, General Manager
Contact Phone/Fax:	630-654-5151
Contact Email:	pkrett@sterigenics.com

SCOPE OF INSPECTION
Compliance with Seal Order issued in 2/15/19

OFF-SITE SURVEILLANCE AND OBSERVATIONS
No on-site semi-truck traffic was observed during the inspection.
Facility Representatives: Chris Lundgren, Technical Training Specialist

IEPA-DIVISION OF RECORDS MANAGEMENT
RELEASABLE

JUL 10 2019

REVIEWER: SAB

INSPECTION FINDINGS

On June 27, 2019, the author conducted an unannounced inspection of Sterigenics in Willowbrook, Illinois. The inspection took place from 2:00 PM to 2:20 PM. The purpose of the inspection was to ensure compliance with the seal order issued on February 15, 2019.

The author met with Chris Lundgren, Technical Training Specialist. Mr. Lundgren provided a tour of the facility (buildings 1 & 2) specifically concentrating on the process and emission units with their associated control equipment.

Facility Operations

Building 1

- All three preconditioning Heat Rooms were empty.
- Sterilization Chambers: All 14 Sterilization Chambers were empty. Observations in the control room showed that all 14 Chambers were off-line/stopped.
- Both Aeration Chambers 1 & 2 were empty
- Both Scrubbers were off-line.
- Ethelene Oxide Barrels
 - 0 Full Drums
 - 0 Partial Drums
 - 0 Empty Drums
- No final product is on-site
- No unprocessed pallets on site

Building 2

- Preconditioning Heat Room was empty
- Sterilization Chambers: All 4 Sterilization Chambers were empty. Observations in the control room showed that all 4 Chambers were off-line/stopped.
- Aeration Chamber was empty
- The AAT Scrubber was off-line
- Ethelene Oxide Barrels
 - 0 Full Drums
 - 0 Partial Drums
 - 0 Empty Drums
- No final product is on-site

Seal Order Verification

All Seal Orders that could be viewed from the main road were still on their respective doors.

Based on observations made on-site, the facility has not been in operation since 2/15/19 and is currently adhering to the Seal Order.

SIGNATURE

REPORT CERTIFICATION:

